



November 21, 2024

William Kotas
Intertek PSI
17 British American Boulevard
Latham, NY 12110

RE: Project: HUDSON FALLS INTERMED SCHOOL
Pace Project No.: 70321592

Dear William Kotas:

Enclosed are the analytical results for sample(s) received by the laboratory on November 07, 2024. The results relate only to the samples included in this report. Results reported herein conform to the applicable TNI/NELAC Standards and the laboratory's Quality Manual, where applicable, unless otherwise noted in the body of the report.

The test results provided in this final report were generated by each of the following laboratories within the Pace Network:

- Pace Analytical Services - Melville

If you have any questions concerning this report, please feel free to contact me.

Sincerely,

Lori A. Beyer
lori.beyer@pacelabs.com
516-370-6014
Project Manager

Enclosures



REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



CERTIFICATIONS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Pace Analytical Services, LLC - Melville, NY

575 Broad Hollow Rd, Melville, NY 11747

Connecticut Certification #: PH-0435

Delaware Certification # NY 10478

Maryland Certification #: 208

Massachusetts Certification #: M-NY026

New Hampshire Certification #: 2987

New Jersey Certification #: NY158

New York Certification #: 10478 Primary Accrediting Body

Pennsylvania Certification #: 68-00350

Rhode Island Certification #: LAO00340

Texas Certification #: T104704582

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

**SAMPLE SUMMARY**

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Lab ID	Sample ID	Matrix	Date Collected	Date Received
70321592001	IN-KT-1	Drinking Water	10/30/24 07:20	11/07/24 07:00
70321592002	IN-KT-2	Drinking Water	10/30/24 07:20	11/07/24 07:00
70321592003	IN-KTB-1	Drinking Water	10/30/24 07:20	11/07/24 07:00
70321592004	IN-NUR -1	Drinking Water	10/30/24 07:20	11/07/24 07:00
70321592005	IN-01-A	Drinking Water	10/30/24 07:20	11/07/24 07:00
70321592006	IN-02-A	Drinking Water	10/30/24 07:20	11/07/24 07:00
70321592007	IN-MFB	Drinking Water	10/30/24 07:25	11/07/24 07:00
70321592008	IN-GFB	Drinking Water	10/30/24 07:25	11/07/24 07:00
70321592009	IN-03-A	Drinking Water	10/30/24 07:30	11/07/24 07:00
70321592010	IN-OFF-1	Drinking Water	10/30/24 07:25	11/07/24 07:00
70321592011	IN-DF-4	Drinking Water	10/30/24 07:30	11/07/24 07:00
70321592012	IN-B-A	Drinking Water	10/30/24 07:35	11/07/24 07:00
70321592013	IN-B-B	Drinking Water	10/30/24 07:35	11/07/24 07:00
70321592014	IN-FAC	Drinking Water	10/30/24 07:40	11/07/24 07:00
70321592015	IN-DF-22	Drinking Water	10/30/24 08:10	11/07/24 07:00
70321592016	IN-DF-25	Drinking Water	10/30/24 08:10	11/07/24 07:00
70321592017	IN-DF-23	Drinking Water	10/30/24 08:10	11/07/24 07:00
70321592018	IN-DF-3	Drinking Water	10/30/24 08:10	11/07/24 07:00
70321592019	IN-03-B	Drinking Water	10/30/24 07:30	11/07/24 07:00
70321592020	IN-NUR-C	Drinking Water	10/30/24 07:20	11/07/24 07:00
70321592021	IN-05-A	Drinking Water	10/30/24 07:40	11/07/24 07:00
70321592022	IN-05-B	Drinking Water	10/30/24 07:40	11/07/24 07:00
70321592023	IN-06-A	Drinking Water	10/30/24 07:45	11/07/24 07:00
70321592024	IN-06-B	Drinking Water	10/30/24 07:45	11/07/24 07:00
70321592025	IN-07-A	Drinking Water	10/30/24 07:50	11/07/24 07:00
70321592026	IN-07-B	Drinking Water	10/30/24 07:50	11/07/24 07:00
70321592027	IN-08	Drinking Water	10/30/24 07:50	11/07/24 07:00
70321592028	IN-09	Drinking Water	10/30/24 07:55	11/07/24 07:00
70321592029	IN-10-A	Drinking Water	10/30/24 07:55	11/07/24 07:00
70321592030	IN-10-B	Drinking Water	10/30/24 07:55	11/07/24 07:00
70321592031	IN-11	Drinking Water	10/30/24 07:55	11/07/24 07:00
70321592032	IN-12	Drinking Water	10/30/24 07:45	11/07/24 07:00
70321592033	IN-13	Drinking Water	10/30/24 07:45	11/07/24 07:00
70321592034	IN-14	Drinking Water	10/30/24 07:45	11/07/24 07:00
70321592035	IN-DF-1-BF	Drinking Water	10/30/24 07:40	11/07/24 07:00
70321592036	IN-15-A	Drinking Water	10/30/24 08:10	11/07/24 07:00
70321592037	IN-16-A	Drinking Water	10/30/24 08:10	11/07/24 07:00

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



SAMPLE SUMMARY

Project: HUDSON FALLS INTERMED SCHOOL
Pace Project No.: 70321592

Lab ID	Sample ID	Matrix	Date Collected	Date Received
70321592038	IN-17-A	Drinking Water	10/30/24 08:10	11/07/24 07:00
70321592039	IN-18-A	Drinking Water	10/30/24 08:10	11/07/24 07:00
70321592040	IN-19	Drinking Water	10/30/24 08:00	11/07/24 07:00
70321592041	IN-20	Drinking Water	10/30/24 08:00	11/07/24 07:00
70321592042	IN-21	Drinking Water	10/30/24 08:00	11/07/24 07:00
70321592043	IN-22	Drinking Water	10/30/24 08:05	11/07/24 07:00
70321592044	IN-23	Drinking Water	10/30/24 08:05	11/07/24 07:00
70321592045	IN-24	Drinking Water	10/30/24 08:05	11/07/24 07:00
70321592046	IN-25	Drinking Water	10/30/24 08:10	11/07/24 07:00
70321592047	IN-26	Drinking Water	10/30/24 08:05	11/07/24 07:00
70321592048	IN-27	Drinking Water	10/30/24 08:15	11/07/24 07:00
70321592049	IN-28-A	Drinking Water	10/30/24 08:00	11/07/24 07:00
70321592050	IN-28-B	Drinking Water	10/30/24 08:00	11/07/24 07:00
70321592051	IN-28-C	Drinking Water	10/30/24 08:00	11/07/24 07:00
70321592052	IN-04-A	Drinking Water	10/30/24 08:00	11/07/24 07:00
70321592053	IN-04-B	Drinking Water	10/30/24 08:00	11/07/24 07:00
70321592054	IN-NUR -2	Drinking Water	10/30/24 07:20	11/07/24 07:00
70321592055	IN-DF-24	Drinking Water	10/30/24 08:10	11/07/24 07:00

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

**SAMPLE ANALYTE COUNT**

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Lab ID	Sample ID	Method	Analysts	Analytes Reported
70321592001	IN-KT-1	EPA 200.8	JP2	1
70321592002	IN-KT-2	EPA 200.8	JP2	1
70321592003	IN-KTB-1	EPA 200.8	JP2	1
70321592004	IN-NUR -1	EPA 200.8	JP2	1
70321592005	IN-01-A	EPA 200.8	JP2	1
70321592006	IN-02-A	EPA 200.8	JP2	1
70321592007	IN-MFB	EPA 200.8	JP2	1
70321592008	IN-GFB	EPA 200.8	JP2	1
70321592009	IN-03-A	EPA 200.8	JP2	1
70321592010	IN-OFF-1	EPA 200.8	JP2	1
70321592011	IN-DF-4	EPA 200.8	JP2	1
70321592012	IN-B-A	EPA 200.8	JP2	1
70321592013	IN-B-B	EPA 200.8	JP2	1
70321592014	IN-FAC	EPA 200.8	JP2	1
70321592015	IN-DF-22	EPA 200.8	JP2	1
70321592016	IN-DF-25	EPA 200.8	JP2	1
70321592017	IN-DF-23	EPA 200.8	JP2	1
70321592018	IN-DF-3	EPA 200.8	JP2	1
70321592019	IN-03-B	EPA 200.8	JP2	1
70321592020	IN-NUR-C	EPA 200.8	JP2	1
70321592021	IN-05-A	EPA 200.8	JP2	1
70321592022	IN-05-B	EPA 200.8	JP2	1
70321592023	IN-06-A	EPA 200.8	JP2	1
70321592024	IN-06-B	EPA 200.8	JP2	1
70321592025	IN-07-A	EPA 200.8	JP2	1
70321592026	IN-07-B	EPA 200.8	JP2	1
70321592027	IN-08	EPA 200.8	JP2	1
70321592028	IN-09	EPA 200.8	JP2	1
70321592029	IN-10-A	EPA 200.8	JP2	1
70321592030	IN-10-B	EPA 200.8	JP2	1
70321592031	IN-11	EPA 200.8	JP2	1
70321592032	IN-12	EPA 200.8	JP2	1
70321592033	IN-13	EPA 200.8	JP2	1
70321592034	IN-14	EPA 200.8	JP2	1
70321592035	IN-DF-1-BF	EPA 200.8	JP2	1
70321592036	IN-15-A	EPA 200.8	JP2	1
70321592037	IN-16-A	EPA 200.8	JP2	1

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



SAMPLE ANALYTE COUNT

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Lab ID	Sample ID	Method	Analysts	Analytes Reported
70321592038	IN-17-A	EPA 200.8	JP2	1
70321592039	IN-18-A	EPA 200.8	JP2	1
70321592040	IN-19	EPA 200.8	JP2	1
70321592041	IN-20	EPA 200.8	JP2	1
70321592042	IN-21	EPA 200.8	JP2	1
70321592043	IN-22	EPA 200.8	JP2	1
70321592044	IN-23	EPA 200.8	JP2	1
70321592045	IN-24	EPA 200.8	JP2	1
70321592046	IN-25	EPA 200.8	JP2	1
70321592047	IN-26	EPA 200.8	JP2	1
70321592048	IN-27	EPA 200.8	JP2	1
70321592049	IN-28-A	EPA 200.8	JP2	1
70321592050	IN-28-B	EPA 200.8	JP2	1
70321592051	IN-28-C	EPA 200.8	JP2	1
70321592052	IN-04-A	EPA 200.8	JP2	1
70321592053	IN-04-B	EPA 200.8	JP2	1
70321592054	IN-NUR -2	EPA 200.8	JP2	1
70321592055	IN-DF-24	EPA 200.8	JP2	1

PACE-MV = Pace Analytical Services - Melville

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-KT-1		Lab ID: 70321592001	Collected: 10/30/24 07:20	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.6	ug/L	1.0	1		11/20/24 18:42	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-KT-2		Lab ID: 70321592002	Collected: 10/30/24 07:20	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.2	ug/L	1.0	1		11/20/24 18:44	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-KTB-1		Lab ID: 70321592003	Collected: 10/30/24 07:20	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.8	ug/L	1.0	1		11/20/24 18:46	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-NUR -1		Lab ID: 70321592004	Collected: 10/30/24 07:20	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.4	ug/L	1.0	1		11/20/24 18:47	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-01-A		Lab ID: 70321592005	Collected: 10/30/24 07:20	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	20.9	ug/L	1.0	1		11/20/24 18:49	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-02-A		Lab ID: 70321592006	Collected: 10/30/24 07:20	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	5.1	ug/L	1.0	1		11/20/24 18:50	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-MFB		Lab ID: 70321592007	Collected: 10/30/24 07:25	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		11/20/24 18:52	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-GFB		Lab ID: 70321592008	Collected: 10/30/24 07:25	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		11/20/24 18:54	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-03-A		Lab ID: 70321592009	Collected: 10/30/24 07:30	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	135	ug/L	1.0	1		11/20/24 18:55	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-OFF-1		Lab ID: 70321592010	Collected: 10/30/24 07:25	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.1	ug/L	1.0	1		11/20/24 19:00	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-DF-4		Lab ID: 70321592011	Collected: 10/30/24 07:30	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.7	ug/L	1.0	1		11/20/24 19:01	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-B-A		Lab ID: 70321592012	Collected: 10/30/24 07:35	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.1	ug/L	1.0	1		11/20/24 19:03	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-B-B		Lab ID: 70321592013	Collected: 10/30/24 07:35	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.5	ug/L	1.0	1		11/20/24 19:04	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-FAC		Lab ID: 70321592014	Collected: 10/30/24 07:40	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		11/20/24 19:06	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-DF-22		Lab ID: 70321592015	Collected: 10/30/24 08:10	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		11/20/24 19:08	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-DF-25		Lab ID: 70321592016	Collected: 10/30/24 08:10	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.4	ug/L	1.0	1		11/20/24 19:09	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-DF-23		Lab ID: 70321592017	Collected: 10/30/24 08:10	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		11/20/24 19:11	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-DF-3		Lab ID: 70321592018	Collected: 10/30/24 08:10	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		11/20/24 19:12	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-03-B		Lab ID: 70321592019	Collected: 10/30/24 07:30	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	9.9	ug/L	1.0	1		11/20/24 19:20	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-NUR-C		Lab ID: 70321592020	Collected: 10/30/24 07:20	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.1	ug/L	1.0	1		11/20/24 19:25	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-05-A		Lab ID: 70321592021	Collected: 10/30/24 07:40	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	15.3	ug/L	1.0	1		11/20/24 19:29	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-05-B		Lab ID: 70321592022	Collected: 10/30/24 07:40	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		11/20/24 19:31	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-06-A	Lab ID: 70321592023	Collected: 10/30/24 07:45	Received: 11/07/24 07:00	Matrix: Drinking Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	2.8	ug/L	1.0	1		11/20/24 19:33	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-06-B		Lab ID: 70321592024	Collected: 10/30/24 07:45	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.3	ug/L	1.0	1		11/20/24 19:37	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-07-A		Lab ID: 70321592025	Collected: 10/30/24 07:50	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.1	ug/L	1.0	1		11/20/24 19:39	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-07-B		Lab ID: 70321592026	Collected: 10/30/24 07:50	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.0	ug/L	1.0	1		11/20/24 19:40	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-08		Lab ID: 70321592027	Collected: 10/30/24 07:50	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.9	ug/L	1.0	1		11/20/24 19:42	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-09		Lab ID: 70321592028	Collected: 10/30/24 07:55	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	2.3	ug/L	1.0	1		11/20/24 19:44	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-10-A		Lab ID: 70321592029	Collected: 10/30/24 07:55	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.2	ug/L	1.0	1		11/20/24 19:45	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-10-B		Lab ID: 70321592030	Collected: 10/30/24 07:55	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.6	ug/L	1.0	1		11/20/24 19:47	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-11	Lab ID: 70321592031	Collected: 10/30/24 07:55	Received: 11/07/24 07:00	Matrix: Drinking Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	4.0	ug/L	1.0	1		11/20/24 19:48	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-12	Lab ID: 70321592032	Collected: 10/30/24 07:45	Received: 11/07/24 07:00	Matrix: Drinking Water					
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville								
Lead	5.6	ug/L	1.0	1		11/20/24 19:50	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-13		Lab ID: 70321592033	Collected: 10/30/24 07:45	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	6.3	ug/L	1.0	1		11/20/24 19:52	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-14		Lab ID: 70321592034	Collected: 10/30/24 07:45	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	3.2	ug/L	1.0	1		11/20/24 19:56	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-DF-1-BF		Lab ID: 70321592035	Collected: 10/30/24 07:40	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		11/20/24 19:58	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-15-A		Lab ID: 70321592036	Collected: 10/30/24 08:10	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		11/20/24 19:59	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-16-A		Lab ID: 70321592037	Collected: 10/30/24 08:10	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		11/20/24 20:01	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-17-A		Lab ID: 70321592038	Collected: 10/30/24 08:10	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		11/20/24 20:03	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-18-A		Lab ID: 70321592039	Collected: 10/30/24 08:10	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		11/20/24 20:07	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-19		Lab ID: 70321592040	Collected: 10/30/24 08:00	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	8.5	ug/L	1.0	1		11/20/24 20:15	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-20		Lab ID: 70321592041	Collected: 10/30/24 08:00	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.2	ug/L	1.0	1		11/20/24 20:19	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-21		Lab ID: 70321592042	Collected: 10/30/24 08:00	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.6	ug/L	1.0	1		11/20/24 20:21	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-22		Lab ID: 70321592043	Collected: 10/30/24 08:05	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		11/20/24 20:22	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-23		Lab ID: 70321592044	Collected: 10/30/24 08:05	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		11/20/24 20:24	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-24		Lab ID: 70321592045	Collected: 10/30/24 08:05	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		11/20/24 20:26	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-25		Lab ID: 70321592046	Collected: 10/30/24 08:10	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		11/20/24 20:27	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-26		Lab ID: 70321592047	Collected: 10/30/24 08:05	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.3	ug/L	1.0	1		11/20/24 20:29	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-27		Lab ID: 70321592048	Collected: 10/30/24 08:15	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		11/20/24 20:34	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-28-A		Lab ID: 70321592049	Collected: 10/30/24 08:00	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		11/20/24 20:35	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-28-B		Lab ID: 70321592050	Collected: 10/30/24 08:00	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		11/20/24 20:37	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-28-C		Lab ID: 70321592051	Collected: 10/30/24 08:00	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		11/20/24 20:38	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-04-A		Lab ID: 70321592052	Collected: 10/30/24 08:00	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.9	ug/L	1.0	1		11/20/24 20:40	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-04-B		Lab ID: 70321592053	Collected: 10/30/24 08:00	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		11/20/24 20:42	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-NUR -2		Lab ID: 70321592054	Collected: 10/30/24 07:20	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.3	ug/L	1.0	1		11/20/24 20:43	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Sample: IN-DF-24		Lab ID: 70321592055	Collected: 10/30/24 08:10	Received: 11/07/24 07:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		11/20/24 20:45	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

QC Batch:	371756	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70321592001, 70321592002, 70321592003, 70321592004, 70321592005, 70321592006, 70321592007, 70321592008, 70321592009, 70321592010, 70321592011, 70321592012, 70321592013, 70321592014, 70321592015, 70321592016, 70321592017, 70321592018		

METHOD BLANK:	1944696	Matrix:	Water
Associated Lab Samples:	70321592001, 70321592002, 70321592003, 70321592004, 70321592005, 70321592006, 70321592007, 70321592008, 70321592009, 70321592010, 70321592011, 70321592012, 70321592013, 70321592014, 70321592015, 70321592016, 70321592017, 70321592018		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	11/20/24 18:27	

LABORATORY CONTROL SAMPLE:	1944697					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	50.9	102	85-115	

MATRIX SPIKE SAMPLE:	1944699						
Parameter	Units	70321591002 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	48.4	97	70-130	

MATRIX SPIKE SAMPLE:	1944701						
Parameter	Units	70321591003 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	50.3	100	70-130	

SAMPLE DUPLICATE:	1944698					
Parameter	Units	70321591002 Result	Dup Result	RPD	Max RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		20	

SAMPLE DUPLICATE:	1944700					
Parameter	Units	70321591003 Result	Dup Result	RPD	Max RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		20	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

QC Batch:	371759	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70321592019, 70321592020, 70321592021, 70321592022, 70321592023, 70321592024, 70321592025, 70321592026, 70321592027, 70321592028, 70321592029, 70321592030, 70321592031, 70321592032, 70321592033, 70321592034, 70321592035, 70321592036, 70321592037, 70321592038		

METHOD BLANK:	1944711	Matrix:	Water
Associated Lab Samples:	70321592019, 70321592020, 70321592021, 70321592022, 70321592023, 70321592024, 70321592025, 70321592026, 70321592027, 70321592028, 70321592029, 70321592030, 70321592031, 70321592032, 70321592033, 70321592034, 70321592035, 70321592036, 70321592037, 70321592038		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	11/20/24 19:14	

LABORATORY CONTROL SAMPLE:	1944712					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	51.9	104	85-115	

MATRIX SPIKE SAMPLE:	1944714						
Parameter	Units	70321592019 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	9.9	50	62.7	106	70-130	

MATRIX SPIKE SAMPLE:	1944716						
Parameter	Units	70321592020 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	2.1	50	56.8	109	70-130	

SAMPLE DUPLICATE:	1944713					
Parameter	Units	70321592019 Result	Dup Result	RPD	Max RPD	Qualifiers
Lead	ug/L	9.9	10	1	20	

SAMPLE DUPLICATE:	1944715					
Parameter	Units	70321592020 Result	Dup Result	RPD	Max RPD	Qualifiers
Lead	ug/L	2.1	2.1	1	20	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

QC Batch:	371760	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70321592039, 70321592040, 70321592041, 70321592042, 70321592043, 70321592044, 70321592045, 70321592046, 70321592047, 70321592048, 70321592049, 70321592050, 70321592051, 70321592052, 70321592053, 70321592054, 70321592055		

METHOD BLANK:	1944726	Matrix:	Water
Associated Lab Samples:	70321592039, 70321592040, 70321592041, 70321592042, 70321592043, 70321592044, 70321592045, 70321592046, 70321592047, 70321592048, 70321592049, 70321592050, 70321592051, 70321592052, 70321592053, 70321592054, 70321592055		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	11/20/24 20:04	

LABORATORY CONTROL SAMPLE:	1944727					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	50.5	101	85-115	

MATRIX SPIKE SAMPLE:	1944729						
Parameter	Units	70321592039 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	52.8	105	70-130	

MATRIX SPIKE SAMPLE:	1944731						
Parameter	Units	70321592040 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	8.5	50	60.2	103	70-130	

SAMPLE DUPLICATE:	1944728					
Parameter	Units	70321592039 Result	Dup Result	RPD	Max RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		20	

SAMPLE DUPLICATE:	1944730					
Parameter	Units	70321592040 Result	Dup Result	RPD	Max RPD	Qualifiers
Lead	ug/L	8.5	8.7	2	20	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALIFIERS

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to dilution of the sample aliquot.

ND - Not Detected at or above adjusted reporting limit.

TNTC - Too Numerous To Count

J - Estimated concentration above the adjusted method detection limit and below the adjusted reporting limit.

MDL - Adjusted Method Detection Limit.

PQL - Practical Quantitation Limit.

RL - Reporting Limit - The lowest concentration value that meets project requirements for quantitative data with known precision and bias for a specific analyte in a specific matrix.

S - Surrogate

1,2-Diphenylhydrazine decomposes to and cannot be separated from Azobenzene using Method 8270. The result for each analyte is a combined concentration.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Indicates the compound was analyzed for, but not detected.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Reported results are not rounded until the final step prior to reporting. Therefore, calculated parameters that are typically reported as "Total" may vary slightly from the sum of the reported component parameters.

Pace Analytical is TNI accredited. Contact your Pace PM for the current list of accredited analytes.

TNI - The NELAC Institute.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
70321592001	IN-KT-1	EPA 200.8	371756		
70321592002	IN-KT-2	EPA 200.8	371756		
70321592003	IN-KTB-1	EPA 200.8	371756		
70321592004	IN-NUR -1	EPA 200.8	371756		
70321592005	IN-01-A	EPA 200.8	371756		
70321592006	IN-02-A	EPA 200.8	371756		
70321592007	IN-MFB	EPA 200.8	371756		
70321592008	IN-GFB	EPA 200.8	371756		
70321592009	IN-03-A	EPA 200.8	371756		
70321592010	IN-OFF-1	EPA 200.8	371756		
70321592011	IN-DF-4	EPA 200.8	371756		
70321592012	IN-B-A	EPA 200.8	371756		
70321592013	IN-B-B	EPA 200.8	371756		
70321592014	IN-FAC	EPA 200.8	371756		
70321592015	IN-DF-22	EPA 200.8	371756		
70321592016	IN-DF-25	EPA 200.8	371756		
70321592017	IN-DF-23	EPA 200.8	371756		
70321592018	IN-DF-3	EPA 200.8	371756		
70321592019	IN-03-B	EPA 200.8	371759		
70321592020	IN-NUR-C	EPA 200.8	371759		
70321592021	IN-05-A	EPA 200.8	371759		
70321592022	IN-05-B	EPA 200.8	371759		
70321592023	IN-06-A	EPA 200.8	371759		
70321592024	IN-06-B	EPA 200.8	371759		
70321592025	IN-07-A	EPA 200.8	371759		
70321592026	IN-07-B	EPA 200.8	371759		
70321592027	IN-08	EPA 200.8	371759		
70321592028	IN-09	EPA 200.8	371759		
70321592029	IN-10-A	EPA 200.8	371759		
70321592030	IN-10-B	EPA 200.8	371759		
70321592031	IN-11	EPA 200.8	371759		
70321592032	IN-12	EPA 200.8	371759		
70321592033	IN-13	EPA 200.8	371759		
70321592034	IN-14	EPA 200.8	371759		
70321592035	IN-DF-1-BF	EPA 200.8	371759		
70321592036	IN-15-A	EPA 200.8	371759		
70321592037	IN-16-A	EPA 200.8	371759		
70321592038	IN-17-A	EPA 200.8	371759		
70321592039	IN-18-A	EPA 200.8	371760		
70321592040	IN-19	EPA 200.8	371760		
70321592041	IN-20	EPA 200.8	371760		
70321592042	IN-21	EPA 200.8	371760		
70321592043	IN-22	EPA 200.8	371760		
70321592044	IN-23	EPA 200.8	371760		
70321592045	IN-24	EPA 200.8	371760		
70321592046	IN-25	EPA 200.8	371760		
70321592047	IN-26	EPA 200.8	371760		
70321592048	IN-27	EPA 200.8	371760		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: HUDSON FALLS INTERMED SCHOOL

Pace Project No.: 70321592

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
70321592049	IN-28-A	EPA 200.8	371760		
70321592050	IN-28-B	EPA 200.8	371760		
70321592051	IN-28-C	EPA 200.8	371760		
70321592052	IN-04-A	EPA 200.8	371760		
70321592053	IN-04-B	EPA 200.8	371760		
70321592054	IN-NUR -2	EPA 200.8	371760		
70321592055	IN-DF-24	EPA 200.8	371760		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



Pace* Location Requested (City/State):
 Pace Analytical Long Island NY
 575 Broad Hollow Rd, Melville, NY 11747

Company Name: Intertek-PSI
 Street Address: 17 British American Blvd, Latham, NY 12210

Customer Project #: 08215514
 Project Name: HUDSON FALLS CENTRAL SCHOOL DISTRICT

Site Collection Info/Facility ID (as applicable):
 Intermediate School

Time Zone Collected: [] AK [] PT [] MT [] CT [] ET [] AT
 Data Deliverables: [] Level II [] Level III [] Level IV [] EQUIS [] Other

Contact/Report To: William Kotas
 Phone #: (518) 377-9841
 E-Mail: william.kotas@intertek.com
 Cc E-Mail:

Invoice To: PSI Latham Accounts Payable
 Invoice E-Mail: LathamAR@intertek.com

Purchase Order # (if applicable):
 Quote #: CR-BOCES RCO #23-057
 County/State origin of sample(s): New York

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

LAB USE ONLY- Affix Workorder/Login Label Here
WO#: 70321592



**Container Size: (1) 1L, (2) 500mL, (3) 250mL, (4) 125mL, (5) 100mL, (6) 40mL, (7) Encore, (8) TerraCore, (9) Other

***Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) NaOH, (6) Zn Acetate, (7) NaHSO4, (8) Sod. Thiosulfate, (9) Ascorbic Acid, (10) MeOH, (11) Other

Specify Container Size **

Identify Container Preservative Type***

Analysis Requested

Proj. Mgr:
 Lori Beyer
 AcctNum / Client ID:
 AR 70-102260
 Table #:
 Profile / Template:
 10367
 Prelog / Bottle Ord. ID:

Customer Sample ID	Matrix *	Comp / Grab	Collected (or Composite Start) Date	Time	Composite End Date	Time	Res. CLZ	Number & Type of Containers Plastic Glass	2008 Drinking Water (Pb only)	Preservation non-compliance identified for sample.
IN-KT-1	DW	G	10/30/24	7:30				1	X	
IN-KT-2			10/30/24	7:30						
IN-KTB-1			10/30/24	7:30						
IN-NUR-1			10/30/24	7:30						
IN-NUR-2			10/30/24	7:30						
IN-01-A			10/30/24	7:30						
IN-02-A			10/30/24	7:30						
IN-MFB			10/30/24	7:35						
IN-GFB			10/30/24	7:35						
IN-03-A			10/30/24	7:30						

* Matrix Codes (insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Regulatory Program (DW, RCRA, etc.) as applicable: NY Lead in School DW

Rush (Pre-approval required):
 [] 12 Day [] 3 day [] 15 day [] Other

Date Results Requested: Standard 10 business day

Field Filtered (if applicable): [] Yes [] No

Additional Instructions from Pace:

Collected By: Richard Paszkiewicz
 Printed Name: Richard Paszkiewicz
 Signature:

Thermometer ID:
 Correction Factor (°C):
 Obs. Temp. (°C):
 Corrected Temp. (°C):

Tracking Number:
 Date/Time: 11/6/24 10:45AM
 Received by/Company: Signature: Richard Paszkiewicz

Date/Time: 11/7/24 9:00
 Delivered by: [] In-Person [] Courier
 [] FedEx [] UPS [] Other

Date/Time:
 Received by/Company: Signature:

Date/Time:
 Received by/Company: Signature:

Date/Time:
 Received by/Company: Signature:

Date/Time:
 Received by/Company: Signature:

Date/Time:
 Received by/Company: Signature:

Date/Time:
 Received by/Company: Signature:

Date/Time:
 Received by/Company: Signature:

Date/Time:
 Received by/Company: Signature:

Date/Time:
 Received by/Company: Signature:

Date/Time:
 Received by/Company: Signature:

Date/Time:
 Received by/Company: Signature:

Date/Time:
 Received by/Company: Signature:

Date/Time:
 Received by/Company: Signature:

Date/Time:
 Received by/Company: Signature:

Date/Time:
 Received by/Company: Signature:

Date/Time:
 Received by/Company: Signature:

Company Name: **Intertek-PSI**
 Street Address: **17 British American Blvd, Latham, NY 12210**

Customer Project #: **08215514**
 Project Name: **HUDSON FALLS CENTRAL SCHOOL DISTRICT**

Site Collection Info/Facility ID (as applicable):
Intermediate School

Time Zone Collected: [] AK [] PT [] MT [] CT [X] ET

Data Deliverables:
 [] Level II [] Level III [] Level IV
 [] EQUIS
 [] Other _____

Regulatory Program (DW, RCRA, etc.) as applicable: **NY Lead In School DW**

Rush (Pre-approval required):
 [] 2 Day [] 3 day [] 5 day [] Other _____
 Date Results Requested: **Standard 10 business day**

Field Filtered (if applicable): [] Yes [] No
 Analyte: _____

Matrix Codes (insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected (or Composite Start)		Res. CLZ	Number & Type of Containers
			Date	Time		
IN-OFF-1	DW	G	10/30/24	7:35		1 Plastic
IN-DF-4			10/30/24	7:30		
IN-B-A			10/30/24	7:35		
IN-B-B			10/30/24	7:35		
IN-FAC			10/30/24	7:40		
IN-DF-22			10/30/24	8:10		
IN-DF-25			10/30/24	8:10		
IN-DF-23			10/30/24	8:10		
IN-DF-24			10/30/24	8:10		
IN-DF-3			10/30/24	8:10		

Customer Remarks / Special Conditions / Possible Hazards:
Lead

Relinquished by/Company: (Signature) *[Signature]*
 Date/Time: **11/6/24 10:45am**
 Relinquished by/Company: (Signature) *[Signature]*
 Date/Time: **11/6/24**
 Relinquished by/Company: (Signature) *[Signature]*
 Date/Time: _____
 Relinquished by/Company: (Signature) *[Signature]*
 Date/Time: _____

Received by/Company: (Signature) *[Signature]*
 Date/Time: **11/6/24 10:45am**
 Received by/Company: (Signature) *[Signature]*
 Date/Time: **11/6/24 9:00**
 Received by/Company: (Signature) *[Signature]*
 Date/Time: _____
 Received by/Company: (Signature) *[Signature]*
 Date/Time: _____

Tracking Number: **1111111111**
 Delivered by: [] In-Person [] Courier
 [] FedEx [] UPS [] Other
 Page: **2** of **6**

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Contact/Report To: **William Kotas**
 Phone #: **(518) 377-9841**
 E-Mail: **william.kotas@intertek.com**
 Cc E-Mail: _____

Invoice To: **PSI Latham Accounts Payable**
 Invoice E-Mail: **LathamAR@intertek.com**

Purchase Order # (if applicable): _____
 Quote #: **CR-BOCES RCO #23-057**

Specify Container Size **
 Identify Container Preservative Type***

Analysis Requested

Preservation non-conformance identified for sample:
 Proj. Mgr: **Lori Beyer**
 ActNum / Client ID: **AR 70-102260**
 Table #: _____
 Profile / Template: **40367**
 Prelog / Bottle Ord. ID: _____
 Sample Comment: _____

Lab Use Only

Additional Instructions from Pace*:
 # Coolers: _____ Thermometer ID: _____ Obs. Temp. (°C) _____ Corrected Temp. (°C) _____



Scan QR Code for instructions

**Container Size: (1) 1L, (2) 500mL, (3) 250mL, (4) 125mL, (5) 100mL, (6) 40mL, (7) EriCore, (8) TerraCore, (9) Other

*** Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) NaOH, (6) Zn Acetate, (7) NaHSO4, (8) Soc. Thiosulfate, (9) Ascorbic Acid, (10) MeOH, (11) Other

CHAIN-OF-CUSTODY Analytical Request Document
Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

LAB USE ONLY - Affix Workorder/Login Label Here



Scan QR Code for instructions

Company Name: Intertek-PSI
Street Address: 17 British American Blvd, Latham, NY 12120
Contact/Report To: William Kotas
Phone #: (518) 377-9841
E-Mail: william.kotas@intertek.com
Cc E-Mail:
Invoice To: PSI Latham Accounts Payable
Invoice E-Mail: LathamAR@intertek.com
Purchase Order # (if applicable):
Quote #: CR-BOCES RCO #23-057

Customer Project #: 08215514
Project Name: HUDSON FALLS CENTRAL SCHOOL DISTRICT
Site Collection Info/Facility ID (as applicable):
Intermediate School

County / State origin of sample(s): New York
Regulatory Program (DW, RCRA, etc.) as applicable: NY Lead in School DW
Rush (Pre-approval required): DW PWSID # or WW Permit # as applicable:
 12 Day 15 day Other _____
Date Results Requested: Standard 10 business day
Field Filtered (if applicable): Yes No
Analysis:

Time Zone Collected: AK MT CT ET
Data Deliverables:
 Level II Level III Level IV
 EQUIS
 Other _____

* Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (O), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SD), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected (or Composite Start)		Composite End		Res. CL2	Number & Type of Containers		Sample Comment
			Date	Time	Date	Time		Plastic	Glass	
IN-03-B	DW	G	10/30/24	7:30				1		
IN-NVR-C			10/30/24	7:30						
IN-05-A			10/30/24	7:40						
IN-05-B			10/30/24	7:40						
IN-06-A			10/30/24	7:45						
IN-06-B			10/30/24	7:45						
IN-07-A			10/30/24	7:50						
IN-07-B			10/30/24	7:50						
IN-08			10/30/24	7:50						
IN-09			10/30/24	7:55						

Additional Instructions from Pace:

Collected By: Richard Paszkiewicz
Printed Name: Richard Paszkiewicz
Signature: _____

Coolers: _____ Thermometer ID: _____ Correction Factor (°C): _____ Obs. Temp. (°C): _____ Corrected Temp. (°C): _____

Date/Time: 11/16/24 10:45am
Date/Time: 11/16/24 10:45am
Date/Time: 11/16/24 10:45am
Date/Time: 11/16/24 10:45am

Received by/Company: (Signature) *Pace*
Received by/Company: (Signature) *Richard Paszkiewicz*
Received by/Company: (Signature)
Received by/Company: (Signature)

Tracking Number: *11724 700*
Delivered by: In-Person Courier
 FedEx UPS Other

Page: *3* of *6*

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

LAB USE ONLY- Affix Workorder/Login Label Here



Scan QR Code for instructions

Company Name: **Intertek-PSI**
 Street Address: **17 British American Blvd, Latham, NY 12210**
 Customer Project #: **08215514**
 Project Name: **HUDSON FALLS CENTRAL SCHOOL DISTRICT**
 Site Collection Info/Facility ID (as applicable): **Intermediate School**

Contact/Report To: **William Kotas**
 Phone #: **(518) 377-9841**
 E-Mail: **william.kotas@intertek.com**
 Cc E-Mail:

Invoice To: **PSI Latham Accounts Payable**
 Invoice E-Mail: **LathamAR@intertek.com**
 Purchase Order # (if applicable):
 Quote #: **CR-BOCES RCO #23-057**

Time Zone Collected: [] AK [] PT [] MT [] CT [] ET [] AT
 Data Deliverables: [] Level II [] Level III [] Level IV [] EQUIS [] Other
 Rush (Pre-approval required): [] 2 Day [] 3 day [] 5 day [] Other
 Date Results Requested: **Standard 10 business day**
 Field Filtered (if applicable): [] Yes [] No
 Analysis:

* Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected (or Composite Start)		Res. CL2	Number & Type of Containers	200g Drinking Water (Pb only)
			Date	Time			
IN-10-A	DW	G	10/30/24	7:55		1	X
IN-10-B			10/30/24	7:55			
IN-11			10/30/24	7:55			
IN-12			10/30/24	7:45			
IN-13			10/30/24	7:45			
IN-14			10/30/24	7:45			
IN-DF-1-BF			10/30/24	7:40			
IN-15-A			10/30/24	8:10			
IN-16-A			10/30/24	8:10			
IN-17-A			10/30/24	8:10			

Customer Remarks / Special Conditions / Possible Hazards:
Lead

Collected By: **Richard Paszkiewicz**
 Printed Name: Richard Paszkiewicz
 Signature: *[Signature]*

Received by/Company: **PSI**
 Date/Time: **11/6/24 10:45am**
 Signature: *[Signature]*

Received by/Company: **PSI**
 Date/Time: **11/7/24 7:00**
 Signature: *[Signature]*

Received by/Company: **PSI**
 Date/Time: **11/6/24 10:45**
 Signature: *[Signature]*

Received by/Company: **PSI**
 Date/Time: **11/7/24 7:00**
 Signature: *[Signature]*

Tracking Number: **116 10:45**
 Delivered by: [] In-Person [] Courier [] FedEx [] UPS [] Other

Date/Time: **11/6/24 10:45**
 Signature: *[Signature]*

Date/Time: **11/7/24 7:00**
 Signature: *[Signature]*

Date/Time: **11/6/24 10:45**
 Signature: *[Signature]*

Date/Time: **11/7/24 7:00**
 Signature: *[Signature]*

4 of 6

Company Name: **Intertek-PSI**
 Street Address: **17 British American Blvd, Latham, NY 12120**

Customer Project #: **08215514**
 Project Name: **HUDSON FALLS CENTRAL SCHOOL DISTRICT**

Site Collection Info/Facility ID (as applicable):
Intermediate School

Time Zone Collected: [] AK [] PT [] MT [] CT [] ET

Data Deliverables:
 [] Level II [] Level III [] Level IV
 [] EQUIS
 [] Other

* Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (O), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

CHAIN-OF-CUSTODY Analytical Request Document
 Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Contact/Report To: **William Kotas**
 Phone #: **(518) 377-9841**
 E-Mail: **william.kotas@intertek.com**
 Cc E-Mail:

Invoice To: **PSI Latham Accounts Payable**
 Invoice E-Mail: **LathamAR@intertek.com**

Purchase Order # (if applicable):
 Quote #: **CR-BOCES RCO #23-057**

County / State origin of sample(s): **New York**

Regulatory Program (DW, RCRA, etc.) as applicable: **NY Lead in School DW**
 DW PWSID # or WW Permit # as applicable:
 Rush (Pre-approval required):
 [] 2 Day [] 3 day [] 5 day [] Other
 Date Results Requested: **Standard 10 business day**

Field Filtered (if applicable): [] Yes [] No
 Analytic:



Scan QR Code for instructions

LAB USE ONLY- Affix Workorder/Login Label Here

Specify Container Size **
 Identify Container Preservative Type ***
 *** Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) NaOH, (6) Zn Acetate, (7) NaHSO4, (8) Sod. Thiosulfate, (9) Ascorbic Acid, (10) MeOH, (11) Other

Analyses Requested
 Lab Use Only
 Proj. Mgr: **Lori Beyer**
 AcctNum / Client ID: **AR 70-102260**
 Table #: **10967**
 Profile / Template:
 Prelog / Bottle Ord. ID:

Customer Sample ID	Matrix *	Comp / Grab	Collected (or Composite Start) Date	Time	Composite End Date	Time	Res. CLZ	Number & Type of Containers	Plastic	Glass
IN-18-A	DW	G	10/30/24	8:10				1		
IN-19			10/30/24	8:20						
IN-20			10/30/24	8:00						
IN-21			10/30/24	8:05						
IN-22			10/30/24	8:05						
IN-23			10/30/24	8:05						
IN-24			10/30/24	8:05						
IN-25			10/30/24	8:10						
IN-26			10/30/24	8:05						
IN-27			10/30/24	8:15						

Additional Instructions from Pace:
 # Coolers: Thermometer ID: Correction Factor (°C): Obs. Temp. (°C) Corrected Temp. (°C)

Tracking Number:
 Date/Time: **11/6/24 10:45 AM**
 Received by/Company: **Rice**
 Date/Time: **11/7/24 7:00**
 Received by/Company: **PAGE 4**
 Date/Time:
 Received by/Company: **PAGE 4**
 Date/Time:
 Received by/Company: **PAGE 4**
 Date/Time:

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Scan QR Code for instructions

Company Name: **Intertek-PSI**
 Street Address: **17 British American Blvd, Latham, NY 12110**

Contact/Report To: **William Kotas**
 Phone #: **(518) 377-9841**
 E-Mail: william.kotas@intertek.com

Customer Project #: **08215514**
 Project Name: **HUDSON FALLS CENTRAL SCHOOL DISTRICT**

Invoice To: **PSI Latham Accounts Payable**
 Invoice E-Mail: LathamAR@intertek.com

Site Collection Info/Facility ID (as applicable):
Intermediate School

Analysis Requested

Time Zone Collected: [] AK [] PT [] MT [] CT [X] ET

Regulatory Program (DW, RCRA, etc.) as applicable: **New York**

County / State origin of sample(s): **New York**

Specimen ID: **IN-28-A**

Matrix: **DW**

Comp / Grab: **G**

Collected (or Composite Start) Date: **10/30/24** Time: **8:00**

Res. CLZ: **1**

Number & Type of Containers: **1**

Field Filtered (if applicable): [] Yes [] No

Date Results Requested: **Standard 10 business day**

Analysis: **200.8 Drinking Water (Pb only)**

Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected (or Composite Start) Date	Time	Res. CLZ	Number & Type of Containers	Field Filtered (if applicable)	Date Results Requested	Analysis
IN-28-A	DW	G	10/30/24	8:00	1	1	[] Yes [] No	Standard 10 business day	200.8 Drinking Water (Pb only)
IN-28-B			10/30/24	8:00					
IN-28-C			10/30/24	8:00					
IN-04-A			10/30/24	8:00					
IN-04-B			10/30/24	8:00					

Customer Remarks / Special Conditions / Possible Hazards:

Additional Instructions from Pace*:

Collected By: **Richard Paszkiewicz**
 Printed Name: **Richard Paszkiewicz**
 Signature: *[Signature]*

Received by/Company: (Signature) **Paul Pace**
 Date/Time: **11/6/24 10:45 AM**

Received by/Company: (Signature) **William Kotas**
 Date/Time: **11/9/24 7:00**

Received by/Company: (Signature)
 Date/Time:

Received by/Company: (Signature)
 Date/Time:

Tracking Number: **1016**

Delivered by: [] In-Person [] Courier [] FedEx [] UPS [] Other

Page: **6** of **6**

MO#: 70321592
PM: LAB
Due Date: 11/21/24
CLIENT: INTER-LATHAM

Project # _____
 Client Name: **Inter-Latham**
 Courier: Fed Ex UPS USPS Client Commercial Other

Tracking #: _____
 Custody Seal on Cooler/Box Present: Yes No
 Seals Intact: Yes No
 Temperature Blank Present: Yes No
 Type of Ice: Wet Blue None

Thermometer Used: **TH11**
 Correction Factor: **0.3+**
 Cooler Temperature (C): **18.6**
 Cooler Temperature Corrected (C): **18.9**
 Date/Time 5035A kits placed in freezer _____
 Temp should be above freezing to 5.0°C

USDA Regulated Soil (N/A, water sample)
 Did samples originate in a quarantine zone within the United States: AL, AR, CA, FL, GA, ID, LA, MS, NC, NM, NY, OK, OR, SC, TN, TX, or VA (check map)? Yes No
 Did samples originate from a foreign source including Hawaii and Puerto Rico? Yes No

If Yes to either question, fill out a Regulated Soil Checklist (ENV-FRM-MELV-0076) and include with SCUR/COC paperwork.
 Date and Initials of person examining contents: **11/13/24 PD**

Sample #	Initial when completed:	Lot # of added preservative:	Date/Time preservative added:
1.			
2.			
3.			
4.			
5.			
6.			
7.			
8.			
9.			
10.			
11.			
12.			

Sample #	Initial when completed:	Lot # of added preservative:	Date/Time preservative added:
13.			
14.			
15.			
16.			
17.			

All containers needing preservation Yes No
 pH paper Lot # **213624**
 All containers needing preservation are found to be in compliance with method recommendation? Yes No
 (HNO₃, H₂SO₄, HCl, NaOH > 9 Sulfide, Yes No
 NAOH > 12 Cyanide) Yes No
 Exceptions: VOA, Coliform, TOC/DOC, Oil and Grease, DRO/8015 (water), Per Method, VOA pH is checked after analysis
 Samples checked for dechlorination: Yes No
 KI starch test strips Lot # _____
 Residual chlorine strips Lot # _____
 SM 4500 CN samples checked for sul Yes No
 Lead Acetate Strips Lot # _____
 HeadSpace in ALK Bottle (>6mm): Yes No
 HeadSpace in VOA Vials (>6mm): Yes No
 Trip Blank Present: Yes No
 Trip Blank Custody Seals Present: Yes No

Client Notification/Resolution: _____
 Field Data Required? Y N
 Person Contacted: _____
 Date/Time: _____
 Comments/Resolution: _____